



valid until: August 5, 2031

Fraunhofer

TESTED[®] DEVICE

STÄUBLI WFT GmbH
Sterimove

Report No. ST 2503-1598

Single product Particle Emission in Cleanroom (atmospheric)

Qualification Certificate

This is to certify that the product mentioned above, provided by

Stäubli WFT GmbH

Sulzbach-Rosenberg, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE
bearing the report number ST 2503-1598.

The mobile pharma robot Sterimove (color: white) was assessed
in compliance with ISO 14644-14. When operated under the
specified test conditions (room temperature: $22^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$;
relative humidity: $45\% \pm 5\%$), it is suitable for use in cleanrooms
fulfilling the specifications of the following Air Cleanliness Classes
according to ISO 14644-1:

Test parameter(s)	Air Cleanliness Class
Movement sequence	4
Charging process	4
Overall result	4

Please note: Transport damages, incorrect installation, oil leakage,
aging behavior, corrosion etc. can influence the test result.

This document only
applies to the named
product in its original state
and is valid for a period of
5 years from the date the
first document was issued.
The document can be
verified under
www.tested-device.com.

Detailed information and
parameters of the test
environment can be found
in the Fraunhofer IPA test
report.

The norms stated generally
refer to the version valid at
the time of the tests.

ST 2503-1598
Report No. first document

--
Report No. current document

Stuttgart, August 5, 2026
Place, date of first document issued

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, head of business unit Testing and Certification



Fraunhofer
IPA