



valid until: October 30, 2030

Fraunhofer

TESTED[®] DEVICE

igus GmbH
C6.29 e-spin stand alone
Report No. IG 2510-1677

Single product
**Particle Emission
in Cleanroom
(atmospheric)**

Qualification Certificate

This is to certify that the product mentioned above, provided by

igus GmbH
Cologne, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE
bearing the report number IG 2510-1677.

The C6.29 e-chain with e-spin without full housing – 8 meter
(color: yellow/gray) was assessed in compliance with
ISO 14644-14. When operated under the specified test conditions
(room temperature: $22^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$; relative humidity: $45\% \pm 5\%$),
it is suitable for use in cleanrooms fulfilling the specifications of
the following Air Cleanliness Classes according to ISO 14644-1:

Test parameter(s)	Air Cleanliness Class
$v_1 = 0.5 \text{ m/s}; a_1 = 1.0 \text{ m/s}^2$	1
$v_2 = 1.0 \text{ m/s}; a_2 = 2.0 \text{ m/s}^2$	3
$v_3 = 2.0 \text{ m/s}; a_3 = 4.0 \text{ m/s}^2$	4
Overall result	4

Please note: Transport damages, incorrect installation, aging
behavior etc. can influence the test result.

IG 2510-1677
Report No. first document

Stuttgart, October 30, 2025
Place, date of first document issued

--
Report No. current document

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, head of business unit Testing and Certification

This document only
applies to the named
product in its original state
and is valid for a period of
5 years from the date the
first document was issued.
The document can be
verified under
www.tested-device.com.

Detailed information and
parameters of the test
environment can be found
in the Fraunhofer IPA test
report.



Fraunhofer
IPA