



valid until: March 3, 2031

Fraunhofer

TESTED[®] DEVICE

PFENNIG Reinigungstechnik GmbH
Ringo GMP Ultra
Report No. PF 2509-1667

Single product
Hygienic Design

DUPLICATE

Qualification Certificate

This is to certify that the product mentioned above, provided by

PFENNIG Reinigungstechnik GmbH
Durach, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number PF 2509-1667.

The flat press Ringo GMP Ultra (color: silver/blue) is principally suitable for use in hygienic areas up to the following GMP Class according to EU GMP Annex 1:

Suitability
up to GMP Class A / B

However, this recommendation only pertains to the operating utility when in a resting state. An overall assessment can only be made after its installation.

This document only applies to the named product in its original state and is valid for a period of 5 years from the date the first document was issued. The document can be verified under www.tested-device.com.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

The norms stated generally refer to the version valid at the time of the tests.

PF 2509-1667 Stuttgart, March 3, 2026
Report No. first document Place, date of first document issued

-- --
Report No. current document Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, head of business unit Testing and Certification