



valid until: November 26, 2030

Fraunhofer
TESTED[®]
DEVICE
Ziehl-Abegg SE
HF C27021 14 black
Report No. ZI 2507-1655

**Statement of
Qualification**

Single product
Outgassing Behavior
Inorganic Acids

Statement of Qualification • Single product

Customer

Ziehl-Abegg SE
Heinz-Ziehl-Strasse
Künzelsau
Germany

Test result / Classification

The outgassing behavior of HF C27021 14 black at the stated temperature was investigated according to ISO 14644-15 and VDI 2083 Part 17. Based on the outgassing rates determined for the specific surfaces, the following material classification was made for the corresponding Contaminant Category:

Tested product

Category: Materials
Subcategory: Plastics
Product name: HF C27021 14 black
(manufacturing date: 5/2025; color: black; article number: 00412286)

Contaminant Category (x)	SER ¹⁾ 23 °C [g/m ² s]	ISO-ACC _m Class (x)
Hydrofluoric acid (HF)	< 2.9 x 10 ⁻⁹	< -8.5
Hydrochloric acid (HCl)	< 2.9 x 10 ⁻⁹	< -8.5
Hydrobromic acid (HBr)	< 2.9 x 10 ⁻⁹	< -8.5
Nitric acid (HNO ₃)	< 2.9 x 10 ⁻⁹	< -8.5
Phosphoric acid (H ₃ PO ₄)	< 2.9 x 10 ⁻⁹	< -8.5
Sulfuric acid (H ₂ SO ₄)	< 2.9 x 10 ⁻⁹	< -8.5

Emission chamber measurements with gas impingement in combination with ion chromatography (IC)

Standards/guidelines: ISO 14644-8, -15; VDI 2083 Part 17
The norms stated generally refer to the version valid at the time of the tests.

Test equipment:

- Measuring station:.....Metrohm Professional IC 850
- Sampling chamber:.....Markes International µCTE

Sample storage:

- Pre-conditioning:
 - Cleanroom Air Cleanliness Class (according to ISO 14644-1):.....ISO 1
 - Airflow velocity:.....0.45 m/s
 - Airflow type:.....vertical laminar flow
 - Temperature:22 °C ± 0.5 °C
 - Relative humidity:45 % ± 5 %
 - Purified air:VOC-filtered

Test procedure parameters:

- Volume of micro emission chamber:45 cm³
- Preconditioning time:> 5 min
- Temperature during emission sampling:23 °C
- Duration of emission sampling:24 h
- Sampling flow rate:100 mL/min

¹⁾The emission rate is calculated using the detected concentration based on the external standard calibration, the analyzed sample surface area or number of samples, the volume of the impingement solution and the sampling duration.

The measuring devices used for the qualification tests are calibrated at regular intervals; their results can be traced back to national and international standards. In cases where no national standards exist, the test procedure implemented complies with the technical regulations and norms applicable at the time of the test. The relevant documentation can be viewed on request at any time.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

Fraunhofer Institute for Manufacturing
Engineering and Automation IPA

ZI 2005-1156
Report No. first document

Stuttgart, May 25, 2020
Place, date of first document issued

Business unit
Testing and Certification

ZI 2507-1655
Report No. current document

Stuttgart, November 26, 2025
Place, current date

Nobelstrasse 12
70569 Stuttgart
Germany

on behalf of 
Dr.-Ing. Frank Bürger, head of business unit Testing and Certification

This document only
applies to the named
product in its original state
and is valid for a period of
5 years from the current
date the document was
issued. The document can
be verified under
www.tested-device.com